

# **MEDICAL DEVICE GUIDANCE DOCUMENT**

## **DEFINITION OF MEDICAL DEVICE**

# Contents

Page

Preface .....	iii
1 Introduction .....	1
2 Purpose .....	1
3 Scope .....	1
4 Terms and definitions .....	1
5 Definition of medical device .....	2

## **Preface**

This Guidance Document was prepared by the Medical Device Authority (MDA) to help the industry and healthcare professionals in their quest to comply with the Medical Device Act (Act 737) and the regulations under it.

This Guidance Document shall be read in conjunction with the current laws and regulations used in Malaysia, which include but not limited to the following-

- a) Medical Device Act 2012 (Act 737); and
- b) Medical Device Regulations 2012.

In this Guidance Document, the following verbal forms are used:

- “shall” indicates a requirement;
- “should” indicates a recommendation;
- “may” indicates a permission; and
- “can” indicates a possibility or a capability.

Irrespective of the requirements of this Guidance Document, MDA has the right to request for information or material, or define conditions not specifically described in this document that is deemed necessary for the purpose of regulatory control.

MDA has put much effort to ensure the accuracy and completeness of this guidance document. In the event of any contradiction between the contents of this document and any written law, the latter should take precedence.

MDA reserves the right to amend any part of the guidance document from time to time.

### **CONTACT INFORMATION**

For further information, please contact:

#### **MEDICAL DEVICE AUTHORITY**

Ministry of Health Malaysia  
Level 6, Prima 9, Prima Avenue II,  
Block 3547, Persiaran APEC,  
63000 Cyberjaya, Selangor,  
MALAYSIA  
Fax: (03) 8230 0200  
Email: [mdb@mdb.gov.my](mailto:mdb@mdb.gov.my)  
Website: <http://www.mdb.gov.my>

## **DEFINITION OF MEDICAL DEVICES 1 Introduction**

The definition of medical device is legally binding in the Medical Device Act 2012 (Act 737) and it determines the scope of regulatory control of the product. This definition is a harmonized definition which adopted from the Global Harmonization Task Force (GHTF) recommendation. This definition also differentiates medical devices from medicinal products which have similar intended purpose by putting the intended primary mode of action in the definition. The intended primary mode of action of medical devices on/in the body is not by means of metabolic, immunological or pharmacological action.

### **2 Purpose**

To provide a guidance regarding the definition of a medical device and information on products which may be considered to be a medical device in the jurisdiction under the Medical Device Act 2012 (Act 737).

### **3 Scope**

This document applies to products that have medical purposes, including those used for the in vitro examination of specimens derived from the human body.

### **4 Terms and definitions**

For the purposes of this document, the following terms and definitions apply.

#### **4.1 Accessory**

An article which is intended specifically by manufacturers to be used together with 'parent' medical device to enable that medical device to achieve its intended purpose or to augment or extend the capabilities of that device in fulfilment of its intended use as a medical device.

#### **4.2 Pharmaceutical**

An interaction between the molecules of the substance in question and a cellular constituent, usually referred to as receptor, which either results in a direct response or which blocks the response to another agent. Although not a completely reliable criterion, the presence of a dose-response correlation is indicative of a pharmacological effect.

#### **4.3 Immunological means**

An action in or on the body by stimulation and/or mobilization of cells and/or products which involved in specific immune reaction.

#### 4.4 Metabolic

An action involves an alteration, including stopping, starting or changing the speed of normal chemical processes participating in and available for normal body function.

### 5 Definition of medical device

Medical device means any instrument, apparatus, implement, machine, appliance, implant, *in vitro* reagent or calibrator, software, material or other similar or related article intended by the manufacturer to be used, alone or in combination, for human beings for the purpose of -

- (i) diagnosis, prevention, monitoring, treatment or alleviation of disease;
- (ii) diagnosis, monitoring, treatment, alleviation of or compensation for an injury;
- (iii) investigation, replacement or modification, or support of the anatomy or of a physiological process;
- (iv) support or sustaining life;
- (v) control of conception;
- (vi) disinfection of medical device; or
- (vii) providing information for medical or diagnostic purpose by means of *in vitro* examination of specimens derived from the human body,

which does not achieve its primary intended action in or on the human body by pharmacological, immunological or metabolic means, but that may be assisted in its intended function by such means.

**NOTE 1:**

*The definition of a device for in vitro examination, for example, reagents, calibrators, sample collection and storage devices, control materials and related instruments or apparatus. The information provided by such an in vitro diagnostic device may be for diagnostic, monitoring or compatibility purpose.*

**NOTE 2:**

*These products may be considered to be medical devices;*

- aids for disabled/handicapped people,*
- accessories for medical devices*
- disinfection substances,*
- devices incorporating animal and human tissues which may meet the requirements of the above definition but are subject to different controls.*

**NOTE 3:**

*Accessories intended specifically by manufacturers to be used together with 'parent' medical device to enable that medical device to achieve its intended purpose should be classified as a medical device in its own right. This may result in the accessory having a different classification than the 'parent' device.*

**NOTE 4:**

*Components are included in the definition of a medical device. Components to medical devices are generally controlled through the manufacturer's quality management system and the conformity assessment procedures for device.*

# **MEDICAL DEVICE AUTHORITY**

## **MINISTRY OF HEALTH, MALAYSIA**

### **Contact Information:**

Registration, Licensing and Enforcement Division  
Medical Device Registration Unit  
Medical Device Authority  
Ministry of Health  
Malaysia

Level 6, Prima 9, Prima Avenue II,  
Block 3547, Persiaran APEC,  
63000 Cyberjaya, Selangor,  
MALAYSIA

**T:** (03) 8230 0300

**F:** (03) 8230 0200

**Website:** <http://www.mdb.gov.my>

